

Lot Number: **B1878**

Reference Material Data

Organism: *Staphylococcus aureus*
Designation: NCTC 10788 (ATCC 6538)
Reference: 56055
BTF Internal Reference: B-SA10788-10K-20
Date of Manufacture: 10 November 2011
Expiry Date: 9 November 2013
Extended Expiry Date#: N/A

Quantification Data

(Method Ref: BB-019)

cfu Count per BioBall[®]

Mean cfu SD

8848.0

543.0

95% Prediction Interval

Lower Limit Upper Limit

7559.8

10136.2

For an explanation of the Prediction Interval see the BioBall Instructions For Use (English version) on the BioBall website - www.bioball.com

Culture Information

No. of Passages: 4
QC Media: Nutrient agar
Plate Incubation Time (hrs): 14 - 24
Plate Incubation Temp °C: 37

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Quality Manager:



Sue Warner

23 November 2011



Accredited for compliance with ISO Guide 34

ACCREDITED FOR
TECHNICAL
COMPETENCE

Accredited Reference Material Producer
Accreditation Number: 14993
Site Number: 15276

Strain sourced from NCTC (National Collection of Type Cultures)

ATCC is a trademark of American Type Culture Collection

*BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall batch has been extended.

Water Industry Enquires:

BTF Pty Ltd.
Sydney Australia
Tel. 61 (0) 2 8877 9150
Fax. 61 (0) 2 8877 9101
www.btfbio.com

Food & Pharmaceutical Industry Enquires:

bioMérieux S.A
69280 Marcy l'Etoile - France
Tel. 33 (0) 4 78 87 20 00
Fax. 33 (0) 4 78 87 20 90
www.biomerieux-industry.com



A BIOMÉRIEUX COMPANY

