

Lot Number: **B1885**

**Reference Material Data**

Organism: *Aspergillus brasiliensis*  
 Designation: NCPF 2275 (ATCC 16404)  
 Reference: 56011  
 BTF Internal Reference: B-AN2275-550-20  
 Date of Manufacture: 21 November 2011  
 Expiry Date: 20 November 2013  
 Extended Expiry Date#: N/A

**Quantification Data**

(Method Ref: BB-019)

**cfu Count per BioBall<sup>®</sup>**

Mean cfu	SD
<b>502.5</b>	<b>38.3</b>

**Quantification Data**

**cfu Count per 100µl aliquot\***

Mean cfu	SD
<b>45.7</b>	<b>7.8</b>
95% Prediction Interval	
Lower Limit	Upper Limit
<b>29.8</b>	<b>61.6</b>

For an explanation of the Prediction Interval see the BioBall Instructions For Use (English version) on the BioBall website - [www.bioball.com](http://www.bioball.com)

**Culture Information**

No. of Passages: 4  
 QC Media: DRBC  
 Plate Incubation Time (hrs): counted at 18, 24 and 48  
 Plate Incubation Temp °C: 37

*Note: Number of passages are indicated from the original strain assuming original strain is zero.*

**Strain Confirmation**

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

**Quality Manager:**



Sue Warner

12 December 2011



Accredited for compliance with ISO Guide 34

Accredited Reference Material Producer  
 Accreditation Number: 14993  
 Site Number: 15276

Strain sourced from NCPF (National Collection of Pathogenic Fungi)

ATCC is a trademark of American Type Culture Collection

\*BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website [www.btfbio.com/cofa.php](http://www.btfbio.com/cofa.php) to determine whether or not the shelf life of your BioBall batch has been extended.

\* QC data using BioBall Re-Hydration Fluid (ref 56021)

**Water Industry Enquires:**

BTF Pty Ltd.  
 Sydney Australia  
 Tel. 61 (0) 2 8877 9150  
 Fax. 61 (0) 2 8877 9101  
[www.btfbio.com](http://www.btfbio.com)



**Food & Pharmaceutical Industry Enquires:**

bioMérieux S.A  
 69280 Marcy l'Etoile - France  
 Tel. 33 (0) 4 78 87 20 00  
 Fax. 33 (0) 4 78 87 20 90  
[www.biomerieux-industry.com](http://www.biomerieux-industry.com)

