

Lot Number: **B1887**

Reference Material Data

Organism: *Bacillus subtilis*
 Designation: NCTC 10400 (ATCC 6633)
 Reference: 56012
 BTF Internal Reference: B-BS10400-550-20
 Date of Manufacture: 23 November 2011
 Expiry Date: 22 November 2013
 Extended Expiry Date[#]: N/A

Quantification Data

(Method Ref: BB-019)

cfu Count per BioBall[®]

Mean cfu	SD
577.9	29.8

Quantification Data

cfu Count per 100µl aliquot*

Mean cfu	SD
52.5	6.6
95% Prediction Interval	
Lower Limit	Upper Limit
39.1	66.0

For an explanation of the Prediction Interval see the BioBall Instructions For Use (English version) on the BioBall website - www.bioball.com

Culture Information

No. of Passages: 4
 QC Media: Nutrient agar
 Plate Incubation Time (hrs): 14 - 24
 Plate Incubation Temp °C: 37

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Quality Manager:



Sue Warner

12 December 2011



Accredited for compliance with ISO Guide 34

Accredited Reference Material Producer
 Accreditation Number: 14993
 Site Number: 15276

Strain sourced from NCTC (National Collection of Type Cultures)

ATCC is a trademark of American Type Culture Collection

[#]BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall batch has been extended.

* QC data using BioBall Re-Hydration Fluid (ref 56021)

Water Industry Enquires:

BTF Pty Ltd.
 Sydney Australia
 Tel. 61 (0) 2 8877 9150
 Fax. 61 (0) 2 8877 9101
www.btfbio.com



Food & Pharmaceutical Industry Enquires:

bioMérieux S.A
 69280 Marcy l'Etoile - France
 Tel. 33 (0) 4 78 87 20 00
 Fax. 33 (0) 4 78 87 20 90
www.biomerieux-industry.com

