

Lot Number: **B1930**

**Reference Material Data**

Organism: *Aspergillus brasiliensis*  
Designation: NCPF 2275 (ATCC 16404)  
Reference: 56022  
BTF Internal Reference: B-AN2275-30-20  
Date of Manufacture: 19 January 2012  
Expiry Date: 18 January 2014  
Extended Expiry Date<sup>#</sup>: N/A

**Quantification Data**

(Method Ref: BB-001)

**cfu Count per BioBall<sup>®</sup>**

Mean cfu	SD
<b>28.1</b>	<b>2.8</b>

95% Prediction Interval

Lower Limit	Upper Limit
<b>22.5</b>	<b>33.7</b>

For an explanation of the Prediction Interval see the BioBall Instructions For Use (English version) on the BioBall website - [www.bioball.com](http://www.bioball.com)

**Culture Information**

No. of Passages: 4  
QC Media: DRBC  
Plate Incubation Time (hrs): counted at 18, 24 and 48  
Plate Incubation Temp °C: 37 Aerobic

*Note: Number of passages are indicated from the original strain assuming original strain is zero.*

**Strain Confirmation**

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

**Quality Officer:**



.....  
Linh Le

17 February 2012



Accredited for compliance with ISO Guide 34

Accredited Reference Material Producer  
Accreditation Number: 14993  
Site Number: 15276

Strain sourced from NCPF (National Collection of Pathogenic Fungi)

ATCC is a trademark of American Type Culture Collection

<sup>#</sup>BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website [www.btfbio.com/cofa.php](http://www.btfbio.com/cofa.php) to determine whether or not the shelf life of your BioBall batch has been extended.

**Water Industry Enquires:**

BTF Pty Ltd.  
Sydney Australia  
Tel. 61 (0) 2 8877 9150  
Fax. 61 (0) 2 8877 9101  
[www.btfbio.com](http://www.btfbio.com)



**Food & Pharmaceutical Industry Enquires:**

bioMérieux S.A  
69280 Marcy l'Etoile - France  
Tel. 33 (0) 4 78 87 20 00  
Fax. 33 (0) 4 78 87 20 90  
[www.biomerieux-industry.com](http://www.biomerieux-industry.com)

