

Lot Number: **B1931**

Reference Material Data

Organism: *Lactobacillus fermentum*
Designation: NCIMB 6991 (ATCC 9338)
Reference: 56054
BTF Internal Reference: B-LF6991-10K-20
Date of Manufacture: 20 January 2012
Expiry Date: 19 January 2014
Extended Expiry Date#: N/A

Quantification Data
(Method Ref: BB-019)

cfu Count per BioBall®	
Mean cfu	SD
7764.0	499.3

95% Prediction Interval	
Lower Limit	Upper Limit
6579.4	8948.6

For an explanation of the Prediction Interval see the BioBall Instructions For Use (English version) on the BioBall website - www.bioball.com

Culture Information


No. of Passages: 4
QC Media: Raka Ray
Plate Incubation Time (hrs): 35 - 44
Plate Incubation Temp °C: 37 Aerobic

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Quality Officer:


.....
Linh Le

13 February 2012



Accredited for compliance with ISO Guide 34
Accredited Reference Material Producer
Accreditation Number: 14993
Site Number: 15276

Strain sourced from NCIMB (National Collection of Industrial, Marine and Food)
ATCC is a trademark of American Type Culture Collection

*BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall batch has been extended.

Water Industry Enquires:

BTF Pty Ltd.
Sydney Australia
Tel. 61 (0) 2 8877 9150
Fax. 61 (0) 2 8877 9101
www.btfbio.com



Food & Pharmaceutical Industry Enquires:

bioMérieux S.A
69280 Marcy l'Etoile - France
Tel. 33 (0) 4 78 87 20 00
Fax. 33 (0) 4 78 87 20 90
www.biomerieux-industry.com

