

Lot Number: **B1959**

**Reference Material Data**

Organism: *Clostridium sporogenes*  
Designation: NCTC 12935 (ATCC 11437)  
Reference: 56004 <sup>10 Vials</sup>  
56014 <sup>20 Vials</sup>  
BTF Internal Reference: B-CS12935-550-10  
B-CS12935-550-20  
Date of Manufacture: 28 February 2012  
Expiry Date: 27 February 2014  
Extended Expiry Date<sup>#</sup>: N/A

**Quantification Data**

(Method Ref: BB-019)

**cfu Count per BioBall<sup>®</sup>**

| Mean cfu     | SD          |
|--------------|-------------|
| <b>525.4</b> | <b>22.4</b> |

**Quantification Data**

**cfu Count per 100µl aliquot\***

| Mean cfu    | SD         |
|-------------|------------|
| <b>47.8</b> | <b>6.3</b> |

95% Prediction Interval

| Lower Limit | Upper Limit |
|-------------|-------------|
| <b>34.9</b> | <b>60.6</b> |

For an explanation of the Prediction Interval see the BioBall Instructions For Use on the BioBall website - [www.bioball.com](http://www.bioball.com)

**Culture Information**

No. of Passages: 4  
QC Media: Horse Blood agar  
Plate Incubation Time (hrs): 40 - 48  
Plate Incubation Temp °C: 37 Anaerobic

Note: Number of passages are indicated from the original strain assuming original strain is zero.

**Strain Confirmation**

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

**Quality Officer:**

.....  
Linh Le

9 May 2012



Accredited for compliance with ISO Guide 34

Accredited Reference Material Producer  
Accreditation Number: 14993  
Site Number: 15276

Strain sourced from NCTC (National Collection of Type Cultures)

ATCC is a trademark of American Type Culture Collection

<sup>#</sup>BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website [www.btfbio.com/cofa.php](http://www.btfbio.com/cofa.php) to determine whether or not the shelf life of your BioBall batch has been extended.

\* QC data using BioBall Re-Hydration Fluid (ref 56021)

**Manufactured by:**

BTF - A bioMérieux Company  
Sydney Australia  
Tel. 61 (0) 2 8877 9150  
Fax. 61 (0) 2 8877 9101  
[www.btfbio.com](http://www.btfbio.com)



**Sales and General enquires:**

bioMérieux S.A  
69280 Marcy l'Etoile - France  
Tel. 33 (0) 4 78 87 20 00  
Fax. 33 (0) 4 78 87 20 90  
[www.biomerieux-industry.com](http://www.biomerieux-industry.com)

