

Lot Number: **B1962**

Reference Material Data

Organism: *Escherichia coli*
Designation: NCTC 12923 (ATCC 8739)
Reference: 56146
BTF Internal Reference: B-EC12923-100M-10
Date of Manufacture: 2 March 2012
Expiry Date: 2 March 2014
Extended Expiry Date[#]: N/A

Quantification Data

(Method Ref: BB-030)

cfu Count per BioBall®

Mean cfu SD

133620000.0 9580280.7

Culture Information

No. of Passages: 4
QC Media: Nutrient agar
Plate Incubation Time (hrs): 14 - 24
Plate Incubation Temp °C: 37 Aerobic

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Quality Manager:



Sue Warner

22 March 2012



Accredited for compliance with ISO Guide 34

Accredited Reference Material Producer
Accreditation Number: 14993
Site Number: 15276

Strain sourced from NCTC (National Collection of Type Cultures)
ATCC is a trademark of American Type Culture Collection

[#]BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall batch has been extended.

Water Industry Enquires:

BTF Pty Ltd.
Sydney Australia
Tel. 61 (0) 2 8877 9150
Fax. 61 (0) 2 8877 9101
www.btfbio.com



Food & Pharmaceutical Industry Enquires:

bioMérieux S.A
69280 Marcy l'Etoile - France
Tel. 33 (0) 4 78 87 20 00
Fax. 33 (0) 4 78 87 20 90
www.biomerieux-industry.com

