

Lot Number: **B1985**

**Reference Material Data**

Organism: *Pseudomonas aeruginosa*  
Designation: NCTC 12924 (ATCC 9027)  
Reference: 56147  
BTF Internal Reference: B-PA12924-100M-10  
Date of Manufacture: 28 March 2012  
Expiry Date: 28 March 2014  
Extended Expiry Date<sup>#</sup>: N/A

**Quantification Data**

(Method Ref: BB-030)

**cfu Count per BioBall®**

Mean cfu                      SD

**106240000.0      6505245.7**

**Culture Information**

No. of Passages: 4  
QC Media: Nutrient agar  
Plate Incubation Time (hrs): 14 - 24  
Plate Incubation Temp °C: 37                      Aerobic

*Note: Number of passages are indicated from the original strain assuming original strain is zero.*

**Strain Confirmation**

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

**Quality Manager:**

Sue Warner

2 May 2012



Accredited for compliance with ISO Guide 34

Accredited Reference Material Producer  
Accreditation Number: 14993  
Site Number: 15276

Strain sourced from NCTC (National Collection of Type Cultures)  
ATCC is a trademark of American Type Culture Collection

<sup>#</sup>BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website [www.btfbio.com/cofa.php](http://www.btfbio.com/cofa.php) to determine whether or not the shelf life of your BioBall batch has been extended.

**Water Industry Enquires:**

BTF Pty Ltd.  
Sydney Australia  
Tel. 61 (0) 2 8877 9150  
Fax. 61 (0) 2 8877 9101  
[www.btfbio.com](http://www.btfbio.com)



**Food & Pharmaceutical Industry Enquires:**

bioMérieux S.A  
69280 Marcy l'Etoile - France  
Tel. 33 (0) 4 78 87 20 00  
Fax. 33 (0) 4 78 87 20 90  
[www.biomerieux-industry.com](http://www.biomerieux-industry.com)

