

Lot Number: **B1992**

**Reference Material Data**

Organism: *Salmonella abony*  
 Designation: ACM 5080 (NCTC 6017)  
 Reference: 56008 10 Vials  
 56018 20 Vials  
 BTF Internal Reference: B-SA5080-550-10  
 B-SA5080-550-20  
 Date of Manufacture: 4 April 2012  
 Expiry Date: 4 April 2014  
 Extended Expiry Date<sup>#</sup>: N/A

**Quantification Data**

(Method Ref: BB-019)

**cfu Count per BioBall<sup>®</sup>**

Mean cfu	SD
<b>550.9</b>	<b>32.1</b>

**Quantification Data**

**cfu Count per 100µl aliquot\***

Mean cfu	SD
<b>50.1</b>	<b>7.6</b>
95% Prediction Interval	
Lower Limit	Upper Limit
<b>34.6</b>	<b>65.6</b>

For an explanation of the Prediction Interval see the BioBall Instructions For Use on the BioBall website - [www.bioball.com](http://www.bioball.com)

**Culture Information**

No. of Passages: 4  
 QC Media: Nutrient agar  
 Plate Incubation Time (hrs): 14 - 24  
 Plate Incubation Temp °C: 37 Aerobic

*Note: Number of passages are indicated from the original strain assuming original strain is zero.*

**Strain Confirmation**

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

**Quality Officer:**



.....  
Linh Le

9 May 2012



Accredited for compliance with ISO Guide 34

Accredited Reference Material Producer  
 Accreditation Number: 14993  
 Site Number: 15276

Strain sourced from ACM (Australian Collection of Microorganisms)

<sup>#</sup>BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website [www.btfbio.com/cofa.php](http://www.btfbio.com/cofa.php) to determine whether or not the shelf life of your BioBall batch has been extended.

\* QC data using BioBall Re-Hydration Fluid (ref 56021)

**Manufactured by:**

BTF - A bioMérieux Company  
 Sydney Australia  
 Tel. 61 (0) 2 8877 9150  
 Fax. 61 (0) 2 8877 9101  
[www.btfbio.com](http://www.btfbio.com)



**Sales and General enquires:**

bioMérieux S.A  
 69280 Marcy l'Etoile - France  
 Tel. 33 (0) 4 78 87 20 00  
 Fax. 33 (0) 4 78 87 20 90  
[www.biomerieux-industry.com](http://www.biomerieux-industry.com)

