

Lot Number: **B2098**

Reference Material Data

Organism: *Burkholderia cepacia*
Designation: NCTC 10743
Reference: 414301 10 Vials
BTF Internal Reference: B-BC10743-30-10
Date of Manufacture: 20 July 2012
Expiry Date: 20 July 2013
Extended Expiry Date#: N/A

Quantification Data

(Method Ref: BB-001)

cfu Count per BioBall®

Mean cfu SD

29.2

3.1

95% Prediction Interval

Lower Limit Upper Limit

22.8

35.6

For an explanation of the Prediction Interval see the BioBall Instructions For Use on the BioBall website - www.bioball.com

Culture Information

No. of Passages: 4
QC Media: Nutrient agar
Plate Incubation Time (hrs): 36 - 60
Plate Incubation Temp °C: 37 Aerobic

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Quality Officer:

.....
Linh Le

9 August 2012

Strain sourced from NCTC (National Collection of Type Cultures)

#BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall batch has been extended.

Manufactured by:

BTF - A bioMérieux Company
Sydney Australia
Tel. 61 (0) 2 8877 9150
Fax. 61 (0) 2 8877 9101
www.btfbio.com



Sales and General enquiries:

bioMérieux S.A
69280 Marcy l'Etoile - France
Tel. 33 (0) 4 78 87 20 00
Fax. 33 (0) 4 78 87 20 90
www.biomerieux-industry.com

