

Lot Number: **B2120**

Reference Material Data

Organism: *Candida albicans*
Designation: NCPF 3179
Reference: 413845 10 Vials
56026 20 Vials
BTF Internal Reference: B-CA3179-30-10
B-CA3179-30-20
Date of Manufacture: 14 August 2012
Expiry Date: 14 August 2014
Extended Expiry Date[#]: N/A

Quantification Data

(Method Ref: BB-001)

cfu Count per BioBall[®]

Mean cfu SD

29.7

2.5

95% Prediction Interval

Lower Limit

Upper Limit

24.7

34.7

For an explanation of the Prediction Interval see the BioBall Instructions For Use on the BioBall website - www.bioball.com

Culture Information

No. of Passages: 4
QC Media: Nutrient agar
Plate Incubation Time (hrs): 40 - 48
Plate Incubation Temp °C: 37 Aerobic

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Quality Officer:

.....
Linh Le

24 September 2012



Accredited for compliance with ISO Guide 34

Accredited Reference Material Producer
Accreditation Number: 14993
Site Number: 15276

Strain sourced from NCPF (National Collection of Pathogenic Fungi)

[#]BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall batch has been extended.

Manufactured by:

BTF - A bioMérieux Company
Sydney Australia
Tel. 61 (0) 2 8877 9150
Fax. 61 (0) 2 8877 9101
www.btfbio.com



Sales and General enquiries:

bioMérieux S.A
69280 Marcy l'Etoile - France
Tel. 33 (0) 4 78 87 20 00
Fax. 33 (0) 4 78 87 20 90
www.biomerieux-industry.com

