

Lot Number: **B2139**

**Reference Material Data**

Organism: *Pseudomonas aeruginosa*  
Designation: NCTC 12924  
Reference: 56007 10 Vials  
56017 20 Vials  
BTF Internal Reference: B-PA12924-550-10  
B-PA12924-550-20  
Date of Manufacture: 5 September 2012  
Expiry Date: 5 September 2014  
Extended Expiry Date<sup>#</sup>: N/A

**Quantification Data**

*(Method Ref: BB-019)*

**cfu Count per BioBall<sup>®</sup>**

Mean cfu	SD
<b>579.3</b>	<b>48.6</b>

**Quantification Data**

**cfu Count per 100µl aliquot\***

Mean cfu	SD
<b>52.7</b>	<b>8.5</b>

95% Prediction Interval

Lower Limit	Upper Limit
<b>35.5</b>	<b>69.8</b>

For an explanation of the Prediction Interval see the BioBall Instructions For Use on the BioBall website - [www.bioball.com](http://www.bioball.com)

**Culture Information**

No. of Passages: 4  
QC Media: Nutrient agar  
Plate Incubation Time (hrs): 14 - 24  
Plate Incubation Temp °C: 37 Aerobic

*Note: Number of passages are indicated from the original strain assuming original strain is zero.*

**Strain Confirmation**

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

**Quality Officer:**

.....  
Linh Le

17 October 2012



Accredited for compliance with ISO Guide 34

Accredited Reference Material Producer  
Accreditation Number: 14993  
Site Number: 15276

Strain sourced from NCTC (National Collection of Type Cultures)

<sup>#</sup>BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website [www.btfbio.com/cofa.php](http://www.btfbio.com/cofa.php) to determine whether or not the shelf life of your BioBall batch has been extended.

\* QC data using BioBall Re-Hydration Fluid (ref 56021)

**Manufactured by:**

BTF - A bioMérieux Company  
Sydney Australia  
Tel. 61 (0) 2 8877 9150  
Fax. 61 (0) 2 8877 9101  
[www.btfbio.com](http://www.btfbio.com)



**Sales and General enquiries:**

bioMérieux S.A  
69280 Marcy l'Etoile - France  
Tel. 33 (0) 4 78 87 20 00  
Fax. 33 (0) 4 78 87 20 90  
[www.biomerieux-industry.com](http://www.biomerieux-industry.com)

