

Lot Number: **B2169**

Reference Material Data

Organism: *Pseudomonas aeruginosa*
Designation: NCTC 12924
Reference: 413831 10 Vials
56144 20 Vials
BTF Internal Reference: B-PA12924-10K-10
B-PA12924-10K-20
Date of Manufacture: 10 October 2012
Expiry Date: 10 October 2014
Extended Expiry Date*: N/A

Quantification Data

(Method Ref: BB-019)

cfu Count per BioBall®

Mean cfu SD

11804.0

1475.3

95% Prediction Interval

Lower Limit Upper Limit

8304.0

15304.0

For an explanation of the Prediction Interval see the BioBall Instructions For Use on the BioBall website - www.bioball.com

Culture Information

No. of Passages: 4
QC Media: Nutrient agar
Plate Incubation Time (hrs): 14 - 24
Plate Incubation Temp °C: 37 Aerobic

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Quality Manager:

Sue Warner

14 November 2012



Accredited for compliance with ISO Guide 34

ACCREDITED FOR
TECHNICAL
COMPETENCE

Accredited Reference Material Producer
Accreditation Number: 14993
Site Number: 15276

Strain sourced from NCTC (National Collection of Type Cultures)

*BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website www.btfbio.com/cofa.php to determine whether or not the shelf life of your BioBall batch has been extended.

Manufactured by:

BTF - A bioMérieux Company
Sydney Australia
Tel. 61 (0) 2 8877 9150
Fax. 61 (0) 2 8877 9101
www.btfbio.com



Sales and General enquiries:

bioMérieux S.A
69280 Marcy l'Etoile - France
Tel. 33 (0) 4 78 87 20 00
Fax. 33 (0) 4 78 87 20 90
www.biomerieux-industry.com

