

Lot Number: **B2229**

Reference Material Data

Organism: *Lactobacillus fermentum*
Designation: NCIMB 6991
Reference: 413830 ^{10 Vials}
56054 ^{20 Vials}
BTF Internal Reference: B-LF6991-10K-10
B-LF6991-10K-20
Date of Manufacture: 6 December 2012
Expiry Date: 6 December 2014
Extended Expiry Date[#]: N/A

Quantification Data

(Method Ref: BB-019)

cfu Count per BioBall®

Mean cfu SD

7644.0

611.3

95% Prediction Interval

Lower Limit Upper Limit

6193.8

9094.2

For an explanation of the Prediction Interval
see the BioBall Instructions For Use on the
BioBall website - www.bioball.com

Culture Information

No. of Passages: 4
QC Media: Raka Ray
Plate Incubation Time (hrs): 40 - 48
Plate Incubation Temp °C: 37 Aerobic

Note: Number of passages are indicated from the original strain assuming original strain is zero.

Strain Confirmation

Bacteria strains are confirmed by genetic typing. Yeast and Mould strains are confirmed by Commonwealth Scientific and Industrial Research Organisation (CSIRO).

Quality Officer:

.....
Linh Le

8 January 2013



Accredited for compliance with ISO Guide 34

Accredited Reference Material Producer
Accreditation Number: 14993
Site Number: 15276

Strain sourced from NCIMB (National Collection of Industrial, Marine and Food Bacteria)

[#]BTF regularly extends the expiry dates on standard BioBall products. Please check the BTF website www.bfbio.com/cofa.php to determine whether or not the shelf life of your BioBall batch has been extended.

Manufactured by:

BTF - A bioMérieux Company
Sydney Australia
Tel. 61 (0) 2 8877 9150
Fax. 61 (0) 2 8877 9101
www.btfbio.com



Sales and General enquiries:

bioMérieux S.A
69280 Marcy l'Etoile - France
Tel. 33 (0) 4 78 87 20 00
Fax. 33 (0) 4 78 87 20 90
www.biomerieux-industry.com

